

**\*\*ABSTRACT NOT FOR CITATION WITHOUT AUTHOR PERMISSION.** The title, authors, and abstract for this completion report are provided below. For a copy of the full completion report, please contact the author via e-mail at [wadecavender@utah.gov](mailto:wadecavender@utah.gov). Questions? Contact the GLFC via email at [frp@glfc.org](mailto:frp@glfc.org).

## **Movements and survival of hatchery reared juvenile cisco (*Coregonus artedii*) in Saginaw Bay, Lake Huron**

**Wade Cavender<sup>1\*</sup>, John Drennan<sup>2</sup>, Gavin Glenney<sup>3</sup>, Anna Forest<sup>1</sup>, Sascha Hallett<sup>4</sup>, Alison Aceves Johnson<sup>1</sup>, Christopher Knupp<sup>5</sup>, Thomas Loch<sup>5</sup>, Joe Marcino<sup>6</sup>, Aimee Reed<sup>7</sup>, Robert Shields<sup>1</sup>, Christine Swan<sup>1</sup>, Danielle VanVliet<sup>1</sup> Skylar Wolf<sup>1</sup>**

- <sup>1</sup> Aquatic Animal Health and Research Center, Utah Division of Wildlife, Logan, Utah
- <sup>2</sup> Fish Health Laboratory, Colorado Parks and Wildlife, Brush, Colorado
- <sup>3</sup> Lamar Fish Health Center, U.S. Fish and Wildlife Service, Lamar, Pennsylvania
- <sup>4</sup> Department of Microbiology, Oregon State University, Corvallis, Oregon
- <sup>5</sup> Department of Fisheries and Wildlife, Michigan State University, East Lansing, Michigan
- <sup>6</sup> Fish Health Laboratory, Arizona Game and Fish Department, Phoenix, Arizona
- <sup>7</sup> Fish Health Services, Oregon Department of Fish and Wildlife, Corvallis, Oregon

*June 2023*

### **ABSTRACT:**

Accurate detection of *Myxobolus cerebralis*, the causative agent of Whirling Disease in salmonids, was compared using three previously developed real-time PCR methods (qPCR). A combination of intra and inter-laboratory analysis using synthetic gBlocks along with naturally and experimentally infected fish tissue was used to evaluate assay performance. Additionally, isolates that represent geographically distinct locations were tested by all three assays. Findings demonstrated that the Kelley 18S (2004) and Cavender 18S (2004) protocols provided high assay sensitivity, repeatability of replicates within and between reactions and reproducible identification of all test samples across multiple laboratories. The Kelley 18S protocol was the only assay that also demonstrated 100% detection accuracy when testing geographically distinct *M. cerebralis* isolates. We determined that the Kelley 18S protocol was robust under variable test conditions (3 thermocyclers, 5 laboratories, multiple operators, and varied sample type), was more sensitive than the Cavender 18S and Hsp70 methods, and provides equivalent detection capabilities to the currently accepted nPCR molecular confirmation method “gold standard” for *M. cerebralis* described in the American Fisheries Society-Fish Health Section Blue Book. Results from this work demonstrate the Kelley 18S qPCR protocol provides a good “fit for purpose” approach for *M. cerebralis* surveillance and confirmation methods.